

# **IMPACT® 326/326M ASPIRATOR SPECIFICATIONS**

<b>Vacuum Range, Continuous:</b>	<b>Adjustable, 0-550 mmHg (0-22" Hg)</b>
<b>Vacuum Range, Intermittent:</b>	<b>Adjustable, 0-200 mmHg (0-8" Hg)</b>
<b>Free Air Flow, Continuous:</b>	<b>30 Liters Per Minute (LPM)</b>
<b>Free Air Flow, Intermittent:</b>	<b>5-8 Liters Per Minute (LPM)</b>
<b>Temperature Operating Range:</b>	<b>-20°C to 49°C (-4°F to 120°F)</b>
<b>Controls:</b>	<b>Power &amp; Mode Selector Switch (Continuous/Power OFF/Intermittent) Vacuum Regulator (Limiter) "ON TIME" Suction Interval Control (adjustable, 5-40 seconds) "OFF TIME" Suction Interval Control (adjustable, 5-40 seconds) Power Source Select Switch (115/230 VAC, located on AC/DC Power Supply)</b>
<b>Displays:</b>	<b>Charge Indicator Lamp External Power Indicator Lamp Battery Charge Indicator (Meter) Vacuum Gauge, Dual-Scale (Metric/English)</b>
<b>Connectors:</b>	<b>Vacuum Inlet Fitting External Power Input Jack</b>
<b>Pump Protection:</b>	<b>Circuit Breaker</b>
<b>Power:</b>	
<b>External 115/230 VAC, 50-400 Hz:</b>	<b>Continuous operation and/or recharge with supplied AC/DC Power Supply</b>
<b>External 12 VDC:</b>	<b>Continuous operation and/or recharge with supplied Auto Power Cable</b>
<b>External 11-30 VDC:</b>	<b>Continuous operation and/or recharge (cable not supplied)</b>
<b>Internal Battery Pack, 12 VDC:</b>	<b>Operating Time, 2-hours (minimum, when cycled as specified)</b>
<b>Recharge Time:</b>	<b>16-hours (maximum)</b>
<b>Internal Battery Pack:</b>	<b>Rechargeable, Sealed Lead Acid</b>
<b>Collection Capacity:</b>	<b>800ml (disposable) x 2 1200ml (disposable) x 2 2000ml (disposable) x 2 1100ml (reusable) x 2</b>
<b>Case Material:</b>	<b>Polycarbonate, color-through, injection-molded</b>

**Case Size:** 24.1cm W x 29.2cm H x 12.4cm D (9.5" W x 11.5" H x 4.87" D)

**Weight:** 5.5 Kg (12 lbs)

**Warranty:** One (1) year, limited, as specified

**Note:** Impact Instrumentation, Inc., reserves the right to change these specifications without prior notification.

p/n 929-0326-00 Rev A Updated 10/11